

KD-1508A Rotary Microtome



Technical Specifications:

Section Thickness Setting Range	1 – 35 μ m for continuous sectioning
The sectioning thickness can be set at any value	> 35 μ m
Maximum Specimen Size	50×40mm
Minimum Setting of Sectioning Thickness	1 μ m
Precision Error	±5%
Dimensions	420×330×330 mm (W×D×H)
Net weight	19kg