

MicroPette plus

121°C FULLY Autoclavable

Features

- Fully autoclavable
- Ergonomic design provides excellent operating experience
- Easy-to-read volume display
- The pipettes cover the volume range from 0.1µL to 5mL
- Easy calibration and maintenance
- Manufactured from innovative materials
- Each MicroPette Plus supplied with an Individual calibration certificate according to ISO8655
- Online calibration is available

MicroPette plus multi-channel

121°C
FULLY
Autoclavable

Features

- Dispensing head rotates for effortless pipetting convenience
- Individual piston and tip cone assembly allow for easy repair and maintenance
- Compound material-made tip cone secures high sealing performance
- Compatible with most universal tip brands
- Online calibration is available





8 channels

Fully autoclavable

The MicroPette Plus pipettes are fully autoclavable, enable easy cleaning and reduce the risk of contamination. Steam autoclaving can be performed at 121°C, 1 bar for 20 minutes. After autoclaving, the pipette must be cooled down and left to dry for 12 hours before use.

It is recommended to check the performance of the pipette after each autoclaving. Greasing and sealing of the pipette piston after every 10th autoclaving will secure an enhanced function.





MicroPette

Features

- 121°C HALF Autoclavable
- Pipettes cover a volume range from 0.1µL to 10mL
- Ergonomic design provides excellent operating experience
- Large display window allows for easy volume identification
- Easy calibration and maintenance
- Each MicroPette supplied with an individual calibration certificate according to ISO8655
- Online calibration is available





MicroPette multi-channel



Features

- 8 and 12 channel pipettes are appropriate for 96 well plates
- Dispensing head rotates for effortless pipetting convenience
- Individual piston and tip cone assemblies allowing easy repair and maintenance
- Compound material-made tip cone secures high sealing performance
- Compatible with most universal tip brands
- Online calibration is available





8 channels

Calibration

All DLAB pipettes have been quality tested according to ISO8655-2:2002 and are supplied with individual calibration certificates. The quality control includes gravimetric testing of each pipette with distilled water at 22°C Our website **www.dlabsci.com** allows users to access DLAB online calibration software and achieve accurate and timely calibration.

Online calibration software is FREE to DLAB pipette users

Specifications

This list is appropriate for MicroPette and MicroPette plus (Adjustable and Fixed volume)

		Siligle-C	illalillei Aujustai	ole Volume Pipett			
Volume Range		Test Volume	Error limits in accordance with ISO8655-2				
	Increment		(Accura	cy error)	(Precision error)		
			%	μL	%	μL	
0.1-2.5µL	0.05µL	2.5µL	2.50%	0.0625	2.00%	0.05	
		1.25µL	3.00%	0.0375	3.00%	0.0375	
		0.25μL	12.00%	0.03	6.00%	0.015	
0.5-10µL	0.1μL	10µL	1.00%	0.1	0.80%	0.08	
		5µL	1.50%	0.075	1.50%	0.075	
		1µL	2.50%	0.025	1.50%	0.015	
2-20µL	0.5μL	20µL	0.90%	0.18	0.40%	0.08	
		10µL	1.20%	0.12	1.00%	0.1	
		2µL	3.00%	0.06	2.00%	0.04	
5-50µL	0.5μL	50µL	0.60%	0.3	0.30%	0.15	
		25µL	0.90%	0.225	0.60%	0.15	
		5µL	2.00%	0.1	2.00%	0.1	
10-100µL	1µL	100µL	0.80%	0.8	0.15%	0.15	
		50µL	1.00%	0.5	0.40%	0.2	
		10µL	3.00%	0.3	1.50%	0.15	
20-200µL	1µL	200μL	0.60%	1.2	0.15%	0.3	
		100μL	0.80%	0.8	0.30%	0.3	
		20µL	3.00%	0.6	1.00%	0.2	
50-200µL	1µL	200μL	0.60%	1.2	0.15%	0.3	
		100µL	0.80%	0.8	0.30%	0.3	
		50µL	1.00%	0.5	0.40%	0.2	
100-1000µL	5µL	1000µL	0.60%	6	0.20%	2	
		500µL	0.70%	3.5	0.25%	1.25	
		100µL	2.00%	2	0.70%	0.7	
200-1000μL	5µL	1000µL	0.60%	6	0.20%	2	
		500µL	0.70%	3.5	0.25%	1.25	
		200μL	0.90%	1.8	0.30%	0.6	
1000-5000µL	50μL	5000µL	0.50%	25	0.15%	7.5	
		2500µL	0.60%	15	0.30%	7.5	
		1000µL	0.70%	7	0.30%	3	
2-10mL	0.1mL	10mL	0.60%	60	0.20%	20	
		5mL	1.20%	60	0.30%	15	
		2mL	3.00%	60	0.60%	12	

DLAB specifications are used as guidelines and the user calibration should refer to the industrial standard ISO 8655

		A			-		
	8-0	channel Adj	ustable Vo	lume Pipet	ttes		
\			Error limits in accordance with ISO8655-2				
Volume Range	Increment	Test Volume	(Accurac	cy error)	(Precisior	n error)	
			%	μL	%	μL	
		10μL	1.50%	0.15	1.50%	0.15	
0.5-10μL	0.1µL	5µL	2.50%	0.125	2.50%	0.125	
		1µL	4.00%	0.04	4.00%	0.04	
		50µL	1.00%	0.5	0.50%	0.25	
5-50µL	0.5μL	25µL	1.50%	0.375	1.00%	0.25	
		5µL	3.00%	0.15	2.00%	0.1	
		300µL	0.70%	2.1	0.25%	0.75	
50-300µL	5µL	150µL	1.00%	1.5	0.50%	0.75	
		50µL	1.50%	0.75	0.80%	0.4	
	12-	channel Adj	justable Vo	olume Pipe	ttes		
Volume	Increment		Error limits in accordance with ISO8655-2				
Range		Test Volume	(Accuracy error)		(Precision error)		
			%	μL	%	μL	
	0.1µL	10µL	1.50%	0.15	1.50%	0.15	
0.5-10μL		5µL	2.50%	0.125	2.50%	0.125	
		1µL	4.00%	0.04	4.00%	0.04	
		50µL	1.00%	0.5	0.50%	0.25	
5-50µL	0.5μL	25µL	1.50%	0.375	1.00%	0.25	
		5µL	3.00%	0.15	2.00%	0.1	
		300µL	0.70%	2.1	0.25%	0.75	
50-300µL	5µL	150µL	1.00%	1.5	0.50%	0.75	
		50μL	1.50%	0.75	0.80%	0.4	
		Fixed '	Volume Pip	oettes			
Volume	Increment	Test Volume	Error limits in accordance with ISO8655-2				
Range			(Accuracy error)		(Precision error)		
			%	μL	%	μL	
5µL	-	5µL	1.3%	0.065	1.2%	0.06	
10µL	-	10μL	0.8%	0.08	0.8%	0.08	
20μL -		20µL	0.6%	0.12	0.5%	0.1	
25µL	-	25µL	0.5%	0.125	0.3%	0.075	
50μL	-	50μL 100μL	0.5%	0.25	0.3%	0.15	
	100μL -		0.5%	0.5	0.3%	0.3	
200µL -		200µL	0.4%	0.8	0.2%	0.4	
250µL -		250µL	0.4%	1.0	0.2%	0.5	
500μL -		500µL	0.3%	1.5	0.2%	1.0	
1000µL		1000µL 2000µL	0.3%	3.0 6.0	0.2% 0.15%	3.0	
2000μL 5000μL		2000μL 5000μL	0.3%	15	0.15%	7.5	
σουσμε	- 1	3000μι	0.570	15	0.1370	7.5	

DLAB specifications are used as guidelines and the user calibration should refer to the industrial standard ISO 8655