

DO ELECTRODE



Main technical indexes

Measurement range	
Dissolved oxygen	(0.0 ~ 20.0)mg/L
Temperature	(0 ~ 40 ° C)
Basic error	
Dissolved oxygen	plus or minus 0.3mg/L
Temperature	+ 1 ° C for
Residual current	not greater than 0.15mg/L
Response time	not more than 30 s (20 ° C at 90% response)
Automatic temperature compensation scope	(0 ~ 40 ° C)